

**INSTITUTE OF TEXTILE TECHNOLOGY,
CHOUDWAR, CUTTACK-754025**

CORRIGENDUM

The Tender Notice No. 1618 dated 15.12.2009 published in the daily news papers i.e. "The Samaj", "The Sambad", "The Dhariritri" & "New Indian Express" and uploaded in the website www.dtetorissa.gov.in may be read with the following provisions:

Items excluded from the tender notice: Dell Computers and Uniline UPS

Items included in the above tender notification: Multi-output DC regulated power supply

Sd/-
Principal(I/C), ITT, Choudwar

SPECIFICATIONS OF MULTI-OUTPUT DC REGULATED POWER SUPPLY

- VOLTAGE : 0-32 V Continuously variable.
- CURRENT : 0-2 A Variable.
- DISPLAY : 3 digits digital display for voltage and current.
- LOAD REGULATOR : 0.05%
- Load current : 5A
- +/-15v selection : Output voltage adjustable from +/-3V to +/-15V DC, with single front panel control, load current 500 mA maximum per output.
- LOAD REGULATION : 0.1%

PROTECTION: For 0-30V sector by means of constant current and for other sections automatic cross over type